

## DATASHEET

# NI 9252

8 AI,  $\pm 10$  V, 24 bit, 50 kS/s/ch Simultaneous



- DSUB or screw terminal connectivity
- 250 V RMS, CAT II, channel-to-earth isolation (screw terminal); 60 V DC, CAT I, channel-to-earth isolation (DSUB)
- $-40$  °C to  $70$  °C operating, 5 g vibration, 50 g shock



**Note** In this document, the NI 9252 with screw terminal and the NI 9252 with DSUB are referred to inclusively as the NI 9252.

The NI 9252 is an 8-channel analog input module for CompactDAQ and CompactRIO systems. Each channel provides a  $\pm 10$  V measurement range, 24-bits of resolution with a maximum sample rate of 50 kS/s. The NI 9252 features numerous programmable hardware filters. By choosing the specific Butterworth and comb filters for your application, you can significantly reduce the noise in the system.

	<b>Kit Contents</b> <ul style="list-style-type: none"><li>• NI 9252</li><li>• NI 9252 Getting Started Guide</li></ul>
	<b>Accessories</b> <ul style="list-style-type: none"><li>• NI 9928 Backshell Connector Kit (Screw Terminal)</li><li>• NI 9923 Screw-Terminal Block (DSUB)</li></ul>

# NI C Series Overview

---



NI provides more than 100 C Series modules for measurement, control, and communication applications. C Series modules can connect to any sensor or bus and allow for high-accuracy measurements that meet the demands of advanced data acquisition and control applications.

- Measurement-specific signal conditioning that connects to an array of sensors and signals
- Isolation options such as bank-to-bank, channel-to-channel, and channel-to-earth ground
- -40 °C to 70 °C temperature range to meet a variety of application and environmental needs
- Hot-swappable

The majority of C Series modules are supported in both CompactRIO and CompactDAQ platforms and you can move modules from one platform to the other with no modification.

## CompactRIO



CompactRIO combines an open-embedded architecture with small size, extreme ruggedness, and C Series modules in a platform powered by the NI LabVIEW reconfigurable I/O (RIO) architecture. Each system contains an FPGA for custom timing, triggering, and processing with a wide array of available modular I/O to meet any embedded application requirement.

## CompactDAQ

CompactDAQ is a portable, rugged data acquisition platform that integrates connectivity, data acquisition, and signal conditioning into modular I/O for directly interfacing to any sensor or signal. Using CompactDAQ with LabVIEW, you can easily customize how you acquire, analyze, visualize, and manage your measurement data.




# Software

## LabVIEW Professional Development System for Windows



- Use advanced software tools for large project development
- Generate code automatically using DAQ Assistant and Instrument I/O Assistant
- Use advanced measurement analysis and digital signal processing
- Take advantage of open connectivity with DLLs, ActiveX, and .NET objects
- Build DLLs, executables, and MSI installers

## NI LabVIEW FPGA Module




- Design FPGA applications for NI RIO hardware
- Program with the same graphical environment used for desktop and real-time applications
- Execute control algorithms with loop rates up to 300 MHz
- Implement custom timing and triggering logic, digital protocols, and DSP algorithms
- Incorporate existing HDL code and third-party IP including Xilinx IP generator functions
- Purchase as part of the LabVIEW Embedded Control and Monitoring Suite

## NI LabVIEW Real-Time Module



- Design deterministic real-time applications with LabVIEW graphical programming
- Download to dedicated NI or third-party hardware for reliable execution and a wide selection of I/O
- Take advantage of built-in PID control, signal processing, and analysis functions
- Automatically take advantage of multicore CPUs or set processor affinity manually
- Take advantage of real-time OS, development and debugging support, and board support
- Purchase individually or as part of a LabVIEW suite

# NI 9252 Circuitry



- Input signals on each channel are buffered, conditioned, and then sampled by an ADC.
- Each AI channel provides an independent signal path and ADC, enabling you to sample all channels simultaneously.
- The module protects each channel from overvoltages.

## Filtering


The NI 9252 uses programmable hardware filtering to provide an accurate representation of in-band signals and reject out-of-band signals. The filters discriminate between signals based on the frequency range, or bandwidth, of the signal.

The NI 9252 programmable hardware filter supports both Butterworth and comb filter responses.


### Butterworth Filter

The NI 9252 has a programmable hardware Butterworth low-pass filter. The Butterworth filter provides two selectable filter orders, each with six selectable cut-off frequencies that are configurable per module. The cut-off frequency ( $f_c$ ) of the filter is independent of the data rate ( $f_s$ ). However, using an external master timebase ( $f_M$ ) will influence both the cut-off frequency ( $f_c$ ) and data rate ( $f_s$ ). The following figures show the overall filter response with different filter settings.

**Figure 1. 4th Order Butterworth Filter Response**




**Figure 2. 2nd Order Butterworth Filter Response**




## Comb Filter

The NI 9252 comb filter frequency response is characterized by deep, evenly spaced notches and an overall roll-off towards higher frequencies. The NI 9252 provides five per module-configurable comb filter settings. The different options provide a trade-off of noise rejection (refer to Idle Channel Noise table) for filter settling time (refer to Settling Time equation) and latency (refer to Input Delay equation). To control the response of the programmable comb filter, you can select to have the first notch at 1, 1/2, 1/4, 1/8 or 1/16 of the data rate. The following figure shows the overall filter response with different filter settings.

**Figure 3. Typical Comb Filter Response**



**Figure 4. Typical Comb Filter Flatness**



## Choosing the Right Filter for your Application

The NI 9252 Butterworth filter response is a low pass filter that allows signals with frequencies below the filter cutoff frequency to pass through while attenuating signals with frequencies higher than the filter cutoff frequency. This is useful to filter out unwanted high frequency noise in a signal. The Butterworth filter has a better flatness in the passband compared to the comb filter.

The NI 9252 Butterworth filter is a programmable-order filter. The different filter orders are characterized by the steepness of the filter response roll-off. The higher the filter order, the steeper the roll-off is. However, the trade-off of using higher order response is the higher input

delay. The NI 9252 Butterworth filter allows user to trade-off between filter roll-off and input delay.

The NI 9252 comb filter frequency response is characterized by deep, evenly spaced notches and an overall roll-off towards higher frequencies. This is useful in rejecting specific frequencies and all its harmonics at a specific data rate. For example, the NI 9252 comb filter rejects powerline frequency of 50 Hz and all its harmonics when running at 50 S/s. The comb filter has lower settling time compared to the Butterworth filter.

For more information about filters, refer to the [Appendix](#).

## Data Rates

The frequency of a master timebase ( $f_M$ ) controls the data rate ( $f_s$ ) of the NI 9252. The NI 9252 includes an internal master timebase with a frequency of 12.8 MHz. Using the internal master timebase of 12.8 MHz results in data rates of 50 kS/s, 33.3333 kS/s, 25 kS/s, 20 kS/s, and so on down to 10 S/s, depending on the decimation rate. However, the data rate must remain within the appropriate data rate range.

The following equation provides the available data rates of the NI 9252:

$$f_s = \frac{f_M}{128 \times a}$$

where  $a$  is the decimation rate.

**Table 1.** Available Data Rates with the Internal Master Timebase

$f_s$ (S/s)	Decimation Rate	$f_s$ (S/s)	Decimation Rate	$f_s$ (S/s)	Decimation Rate
50000.0	2	2272.7	44	347.2	288
33333.3	3	2083.3	48	312.5	320
25000.0	4	2000.0	50	284.1	352
20000.0	5	1785.7	56	260.4	384
16666.7	6	1562.5	64	250.0	400
14285.7	7	1388.9	72	223.2	448
12500.0	8	1250.0	80	200.0	500
11111.1	9	1136.4	88	195.3	512
10000.0	10	1041.7	96	142.1	704
8333.3	12	1000.0	100	125.0	800
7142.9	14	892.9	112	100.0	1000

**Table 1.** Available Data Rates with the Internal Master Timebase (Continued)

$f_s$ (S/s)	Decimation Rate	$f_s$ (S/s)	Decimation Rate	$f_s$ (S/s)	Decimation Rate
6250.0	16	781.3	128	97.7	1024
5555.6	18	694.4	144	60.0 <sup>1</sup>	1666 or 1706 <sup>2</sup>
5000.0	20	625.0	160	50.0 <sup>1</sup>	2000 or 2048 <sup>2</sup>
4545.5	22	568.2	176	10.0 <sup>1</sup>	10000 or 10240 <sup>2</sup>
4166.7	24	520.8	192		
3571.4	28	500.0	200		
3125.0	32	446.4	224		
2777.8	36	400.0	250		
2500.0	40	390.6	256		

The NI 9252 can also accept an external master timebase or export its own master timebase. To synchronize the data rate of an NI 9252 with other modules that use master timebases to control sampling, all of the modules must share a single master timebase source. When using an external timebase with a frequency other than 12.8 MHz, the available data rates of the NI 9252 shift by the ratio of the external timebase frequency to the internal timebase frequency. The programmable filter specifications, expressed in Hz, will also scale with the external timebase. Refer to the software help for information about configuring the master timebase source for the NI 9252.



**Note** The cRIO-9151 R Series Expansion chassis does not support sharing timebases between modules.



**Note** The cRIO-9151 R Series Expansion chassis has different maximum data rates from the CompactRIO and CompactDAQ chassis. Refer to the *Input Characteristics* section for detailed information.

## NI 9252 Specifications

The following specifications are typical for the range -40 °C to 70 °C unless otherwise noted.



**Caution** Observe all instructions and cautions in the user documentation. Using the model in a manner not specified can damage the model and compromise the built-in safety protection. Return damaged models to NI for repair.



**Attention** Suivez toutes les instructions et respectez toutes les mises en garde de la documentation utilisateur. L'utilisation d'un modèle de toute autre façon que celle

<sup>1</sup> When using an external timebase of 13.1072 MHz, this data rate does not change with the ratio of the external to internal clocks.

<sup>2</sup> When using an external timebase of 13.1072 MHz.

spécifiée risque de l'endommager et de compromettre la protection de sécurité intégrée. Renvoyez les modèles endommagés à NI pour réparation.

## Input Characteristics

Number of channels	8 analog input channels
ADC resolution	24 bits
Type of ADC	Delta-Sigma with analog prefiltering
Sampling mode	Simultaneous
Internal master timebase ( $f_M$ )	
Frequency	12.8 MHz
Accuracy	±50 ppm maximum
CompactRIO & CompactDAQ chassis data rate range ( $f_s$ )	
Using internal master timebase	
Minimum	10 S/s
Maximum	50 kS/s
Using external master timebase	
Minimum	0.78 S/s
Maximum	51.367 kS/s
R Series Expansion chassis data rate range ( $f_s$ )	
Using internal master timebase	
Minimum	10 S/s
Maximum	25 kS/s
Data rate	$f_s = \frac{f_M}{128 \times a}$
Overvoltage protection <sup>3</sup>	±30 V
Input resistance (AI <sub>x</sub> to COM)	>10 GΩ
Input voltage range (Differential)	
Minimum	10.50 V
Typical	10.58 V

<sup>3</sup> Up to 4 channels simultaneously

Scaling coefficients	1,261,244 pV/LSB
Maximum input voltage (AI <sub>x</sub> to COM)	±10.5 V
Butterworth filter	
Filter order	2nd or 4th order
Cut-off frequencies <sup>4</sup>	$\frac{f_c \times f_M}{12.8 \text{ MHz}}$
Flatness <sup>5</sup>	$\frac{f_F \times f_M}{12.8 \text{ MHz}}$
Input delay <sup>6</sup>	$(t_D - 2.17 \mu\text{s}) \times \left(\frac{12.8 \text{ MHz}}{f_M}\right) + 2.17 \mu\text{s}$
Input delay tolerance	±100 ns

**Table 2.** Butterworth Filter Cut-off Frequencies and Flatness

Master Timebase Clock (f <sub>M</sub> )	Cut-off Frequencies (f <sub>c</sub> )	2nd Order		4th Order	
		0.1% Flatness (f <sub>F</sub> ) at 0.0087 dB	1% Flatness (f <sub>F</sub> ) at 0.087 dB	0.1% Flatness (f <sub>F</sub> ) at 0.0087 dB	1% Flatness (f <sub>F</sub> ) at 0.087 dB
12.8 MHz	4000 Hz	740 Hz	1445 Hz	1125 Hz	2295 Hz
	2000 Hz	415 Hz	750 Hz	875 Hz	1210 Hz
	1000 Hz	215 Hz	380 Hz	430 Hz	615 Hz
	500 Hz	105 Hz	190 Hz	225 Hz	305 Hz
	250 Hz	55 Hz	95 Hz	115 Hz	155 Hz
	125 Hz	25 Hz	45 Hz	60 Hz	75 Hz



**Note** The specifications in [Table 2](#), on page 10 scale linearly with the master timebase frequency as indicated by the formulas shown in the [Butterworth filter](#) section. For example, on a 2nd Order Butterworth filter, for a master timebase clock of 13.1072 MHz, the cut-off frequency is 4096 Hz and 757.7 Hz of 0.1% Flatness instead of the cut-off frequency of 4000 Hz and 740 Hz of 0.1% Flatness at the 12.8 MHz default internal master timebase clock.

<sup>4</sup> Refer to [Table 2](#), on page 10 for the values of f<sub>c</sub> and f<sub>M</sub>.

<sup>5</sup> Refer to [Table 2](#), on page 10 for the values of f<sub>F</sub> and f<sub>M</sub>.

<sup>6</sup> Refer to [Table 3](#), on page 11 for the values of t<sub>D</sub> and f<sub>M</sub>.


**Table 3.** Butterworth Filter Input Delay

Master Timebase Clock ( $f_M$ )	Cut-off Frequencies ( $f_c$ )	2nd Order		4th Order	
		DC Delay ( $t_D$ )	Maximum Delay ( $t_D$ )	DC Delay ( $t_D$ )	Maximum Delay ( $t_D$ )
12.8 MHz	4000 Hz	98.0 $\mu$ s	104.6 $\mu$ s	136.1 $\mu$ s	158.0 $\mu$ s
	2000 Hz	153.4 $\mu$ s	166.9 $\mu$ s	238.7 $\mu$ s	282.6 $\mu$ s
	1000 Hz	266.2 $\mu$ s	292.9 $\mu$ s	449.1 $\mu$ s	538.8 $\mu$ s
	500 Hz	491.2 $\mu$ s	544.4 $\mu$ s	861.5 $\mu$ s	1038.0 $\mu$ s
	250 Hz	941.3 $\mu$ s	1047.7 $\mu$ s	1700.2 $\mu$ s	2059.7 $\mu$ s
	125 Hz	1841.5 $\mu$ s	2054.2 $\mu$ s	3346.9 $\mu$ s	4055.4 $\mu$ s




**Note** The specifications in [Table 3](#), on page 11 scale with the master timebase frequency as indicated by the formulas shown in the [Butterworth filter](#) section. For example, for a master timebase clock of 13.1072 MHz, the 2nd order Butterworth filter with a 4096 Hz cut-off will have a 95.754  $\mu$ s input DC delay.


**Figure 5.** Butterworth Filter Input Delay (4th Order, with 12.8 MHz Timebase, 4000 Hz, 2000 Hz, 1000 Hz)




**Figure 6.** Butterworth Filter Input Delay (4th Order, with 12.8 MHz Timebase, 500 Hz, 250 Hz, 125 Hz)



**Figure 7.** Butterworth Filter Input Delay (2nd Order, with 12.8 MHz Timebase, 4000 Hz, 2000 Hz, 1000 Hz)



**Figure 8.** Butterworth Filter Input Delay (2nd Order, with 12.8 MHz Timebase, 500 Hz, 250 Hz, 125 Hz)



Comb filter

Programmable first notch  $f_s, f_s/2, f_s/4, f_s/8, f_s/16$

Input delay with comb filter<sup>7</sup>

$$\frac{(A + B)}{f_s} + 2.17\mu s$$

Settling time with comb filter<sup>7</sup>

$$\frac{2(A + B)}{f_s} + 2.17\mu s$$

**Table 4.** Input Delay with Comb Filter

Variable	Value
A	2.4 for $f_s = 50000$
	1.8 for $f_s = 14285.71$ to $33333.33$
	1 for $f_s = 2777.78$ to $12500$
	0.6 for $f_s =$ all other output data rates

<sup>7</sup> Refer to [Table 4](#), on page 13 for the values of A and B.

**Table 4. Input Delay with Comb Filter (Continued)**

Variable	Value
B	0 for filter first notch at $f_s$
	0.5 for filter first notch at $f_s/2$
	1.5 for filter first notch at $f_s/4$
	3.5 for filter first notch at $f_s/8$
	7.5 for filter first notch at $f_s/16$

**Table 5. DC Accuracy**

Measurement Conditions	Percent of Reading <sup>8</sup> (Gain Error)	Percent of Range <sup>9</sup> (Offset Error)
Maximum (-40 °C to 70 °C)	±0.22%	±0.08%
Typical (23 °C, ±5 °C)	±0.06%	±0.01%

Non-linearity	5 ppm
Stability of Accuracy	
Gain drift <sup>8</sup>	7.2 ppm/°C
Offset drift	6.4 $\mu$ V/°C
Passband, -3 dB	Refer to the -3 dB graphs in the <a href="#">Filtering</a> section
Delay linearity ( $f_{in} \leq 24.9$ kHz)	11.16 ns maximum
Channel-to-channel mismatch ( $f_{in} \leq 24.9$ kHz)	
Gain	0.2 dB maximum
Delay	166.67 ns/kHz maximum
Module-to-module mismatch ( $f_{in} \leq 24.9$ kHz)	
Delay	$166.67 \text{ ns/kHz} + \frac{1}{f_M}$
Attenuation @ 2 x oversample rate (23° C)	110 dB

<sup>8</sup> Includes the expected difference in measurement between using single-ended and differential sources due to finite CMRR

<sup>9</sup> Range equals 10.58 V

## Idle Channel Noise

Comb filter with first notch at $f_s$	
$f_s = 50 \text{ kS/s}$	27.2 $\mu\text{VRMS}$
$f_s = 10 \text{ kS/s}$	14.7 $\mu\text{VRMS}$
$f_s = 1 \text{ kS/s}$	5.5 $\mu\text{VRMS}$
$f_s \leq 250 \text{ S/s}$	3.7 $\mu\text{VRMS}$
Butterworth filter, $f_s = 50 \text{ kS/s}$	
$f_c = 4 \text{ kHz}$	14.5 $\mu\text{VRMS}$
$f_c = 1 \text{ kHz}$	7.9 $\mu\text{VRMS}$
$f_c = 125 \text{ Hz}$	3.7 $\mu\text{VRMS}$



**Note** The noise specifications assume the NI 9252 is using the internal master timebase frequency of 12.8 MHz.

## Crosstalk (CH to CH)

NI 9252 with screw terminal	
$f_{in} \leq 1 \text{ kHz}$	120 dB
$f_{in} \leq 5 \text{ kHz}$	105 dB
$f_{in} \leq 20 \text{ kHz}$	80 dB
NI 9252 with DSUB	
$f_{in} \leq 1 \text{ kHz}$	115 dB
$f_{in} \leq 5 \text{ kHz}$	100 dB
$f_{in} \leq 20 \text{ kHz}$	80 dB

## Common mode rejection ratio (CMRR) to COM

$f_{in} \leq 60 \text{ Hz}$	74 dB typical, 69 dB minimum
-----------------------------	------------------------------

## Common mode rejection ratio (CMRR) to Earth Ground

$f_{in} \leq 60 \text{ Hz}$	120 dB minimum
-----------------------------	----------------

## Normal mode rejection ratio (NMRR) using internal or external master timebase of 12.8 MHz<sup>10</sup>

60 S/s, $f_{in} = 60 \text{ Hz} \pm 1 \text{ Hz}$	35 dB minimum
50 S/s, $f_{in} = 50 \text{ Hz} \pm 1 \text{ Hz}$	33 dB minimum
10 S/s, $f_{in} = 50 \text{ Hz}/60 \text{ Hz} \pm 1 \text{ Hz}$	33 dB minimum

<sup>10</sup> Only applicable for comb filter.

Normal mode rejection ratio (NMRR) using external master timebase of 13.1072 MHz<sup>10</sup>

60 S/s, $f_{in} = 60 \text{ Hz} \pm 1 \text{ Hz}$	34 dB minimum
50 S/s, $f_{in} = 50 \text{ Hz} \pm 1 \text{ Hz}$	33 dB minimum
10 S/s, $f_{in} = 50 \text{ Hz}/60 \text{ Hz} \pm 1 \text{ Hz}$	33 dB minimum

## Power Requirements

Power consumption from chassis

Active mode	0.95 W maximum
Sleep mode	91 $\mu\text{W}$ maximum

Thermal dissipation (at 70 °C)

Active mode	1.4 W maximum
Sleep mode	673 mW maximum

## Physical Characteristics

Screw-terminal wiring

Gauge	0.05 mm <sup>2</sup> to 0.82 mm <sup>2</sup> (30 AWG to 18 AWG) copper conductor wire
Wire strip length	5 mm to 6 mm (0.20 in. to 0.24 in.) of insulation stripped from the end
Temperature rating	90 °C, minimum
Torque for screw terminals	0.20 N · m to 0.25 N · m (1.8 lb · in. to 2.2 lb · in.)
Wires per screw terminal	One wire per screw terminal; two wires per screw terminal using a 2-wire ferrule
Ferrules	0.25 mm <sup>2</sup> to 1.0 mm <sup>2</sup>
Connector securement	
Securement type	Screw flanges provided
Torque for screw flanges	0.3 N · m to 0.4 N · m (2.7 lb · in. to 3.5 lb · in.)
Weight	
NI 9252 with screw terminal	134 g (4.7 oz)
NI 9252 with DSUB	149 g (5.3 oz)

# NI 9252 with Screw Terminal Safety Voltages

Connect only voltages that are within the following limits:

Channel-to-channel isolation	None
Channel-to-earth ground isolation	
Up to 3,000 m altitude	
Continuous	250 V RMS, Measurement Category II
Withstand	3,000 V RMS, verified by a 5 s dielectric withstand test
Up to 5,000 m altitude	
Continuous	60 V DC, Measurement Category I
Withstand	1,000 V RMS, verified by a 5 s dielectric withstand test
Overvoltage protection	$\pm 30$ V, between any two pins of the connector <sup>11</sup>



**Caution** When using the NI 9252 with screw terminal above 3,000 m or in explosive atmospheres, do not connect the NI 9252 with screw terminal to signals or use for measurements within Measurement Categories II, III, or IV.



**Attention** Lorsque vous utilisez le NI 9252 with screw terminal à une altitude supérieure à 3 000 m ou dans des atmosphères explosibles, ne le connectez pas à des signaux et ne l'utilisez pas pour effectuer des mesures dans les catégories de mesure II, III ou IV.



**Caution** Do not connect the NI 9252 with screw terminal to signals or use for measurements within Measurement Categories III or IV.



**Attention** Ne connectez pas le NI 9252 with screw terminal à des signaux et ne l'utilisez pas pour effectuer des mesures dans les catégories de mesure III ou IV.

Measurement Category II is for measurements performed on circuits directly connected to the electrical distribution system. This category refers to local-level electrical distribution, such as that provided by a standard wall outlet, for example, 115 V for U.S. or 230 V for Europe.

## NI 9252 with DSUB Isolation Voltages

Channel-to-channel	None
--------------------	------

<sup>11</sup> Up to 4 channels simultaneously.

## Channel-to-earth ground isolation

Up to 3,000 m altitude	
Continuous	60 V DC, Measurement Category I
Withstand	1,000 V RMS, verified by a 5 s dielectric withstand test
Up to 5,000 m altitude	
Continuous	60 V DC, Measurement Category I
Withstand	860 V RMS



**Caution** Do not connect the NI 9252 with DSUB to signals or use for measurements within Measurement Categories II, III, or IV.



**Attention** Ne connectez pas le NI 9252 with DSUB à des signaux et ne l'utilisez pas pour effectuer des mesures dans les catégories de mesure II, III ou IV.

Measurement Category I is for measurements performed on circuits not directly connected to the electrical distribution system referred to as *MAINS* voltage. MAINS is a hazardous live electrical supply system that powers equipment. This category is for measurements of voltages from specially protected secondary circuits. Such voltage measurements include signal levels, special equipment, limited-energy parts of equipment, circuits powered by regulated low-voltage sources, and electronics.



**Note** Measurement Categories CAT I and CAT O are equivalent. These test and measurement circuits are for other circuits not intended for direct connection to the MAINS building installations of Measurement Categories CAT II, CAT III, or CAT IV.

## Hazardous Locations

U.S. (UL)	Class I, Division 2, Groups A, B, C, D, T4; Class I, Zone 2, AEx nA IIC T4 Gc
Canada (C-UL)	Class I, Division 2, Groups A, B, C, D, T4; Ex nA IIC T4 Gc
Europe (ATEX) and International (IECEX)	Ex nA IIC T4 Gc DEMKO 12 ATEX 1202658X IECEX UL 14.0089X

# Safety Compliance and Hazardous Locations Standards

This product is designed to meet the requirements of the following electrical equipment safety standards for measurement, control, and laboratory use:

- IEC 61010-1, EN 61010-1
- UL 61010-1, CSA C22.2 No. 61010-1
- EN 60079-0, EN 60079-15
- IEC 60079-0: Ed 6, IEC 60079-15; Ed 4
- UL 60079-0; Ed 6, UL 60079-15; Ed 4
- CSA C22.2 No. 60079-0, CSA C22.2 No. 60079-15



**Note** For UL and other safety certifications, refer to the product label or the [Product Certifications and Declarations](#) section.

## Electromagnetic Compatibility Standards

This product meets the requirements of the following EMC standards for sensitive electrical equipment for measurement, control, and laboratory use:

- EN 61326-1 (IEC 61326-1): Class A emissions; Industrial immunity
- EN 55011 (CISPR 11): Group 1, Class A emissions
- AS/NZS CISPR 11: Group 1, Class A emissions
- FCC 47 CFR Part 15B: Class A emissions
- ICES-001: Class A emissions



**Note** Group 1 equipment (per CISPR 11) is any industrial, scientific, or medical equipment that does not intentionally generate radio frequency energy for the treatment of material or inspection/analysis purposes.



**Note** In the United States (per FCC 47 CFR), Class A equipment is intended for use in commercial, light-industrial, and heavy-industrial locations. In Europe, Canada, Australia and New Zealand (per CISPR 11) Class A equipment is intended for use only in heavy-industrial locations.



**Note** For EMC declarations and certifications, and additional information, refer to the [Online Product Certification](#) section.



**Notice** Conducted RF interference on the I/O ports of the NI 9252 can adversely affect its measurement accuracy.

# CE Compliance

This product meets the essential requirements of applicable European Directives, as follows:

- 2014/35/EU; Low-Voltage Directive (safety)
- 2014/30/EU; Electromagnetic Compatibility Directive (EMC)
- 2014/34/EU; Potentially Explosive Atmospheres (ATEX)

## Product Certifications and Declarations

Refer to the product Declaration of Conformity (DoC) for additional regulatory compliance information. To obtain product certifications and the DoC for NI products, visit [ni.com/product-certifications](http://ni.com/product-certifications), search by model number, and click the appropriate link.

## Shock and Vibration

To meet these specifications, you must panel mount the system.

### Operating vibration

Random	5 g RMS, 10 Hz to 500 Hz
Sinusoidal	5 g, 10 Hz to 500 Hz
Operating shock	30 g, 11 ms half sine; 50 g, 3 ms half sine; 18 shocks at 6 orientations

## Environmental

Refer to the manual for the chassis you are using for more information about meeting these specifications.

Operating temperature (IEC 60068-2-1, IEC 60068-2-2)	-40 °C to 70 °C
Storage temperature (IEC 60068-2-1, IEC 60068-2-2)	-40 °C to 85 °C
Ingress protection	IP40
Operating humidity (IEC 60068-2-30)	10% RH to 90% RH, noncondensing
Storage humidity (IEC 60068-2-30)	5% RH to 95% RH, noncondensing
Pollution Degree	2
Maximum altitude	5,000 m

Indoor use only.

## Environmental Management

NI is committed to designing and manufacturing products in an environmentally responsible manner. NI recognizes that eliminating certain hazardous substances from our products is beneficial to the environment and to NI customers.

For additional environmental information, refer to the *Commitment to the Environment* web page at [ni.com/environment](http://ni.com/environment). This page contains the environmental regulations and directives with which NI complies, as well as other environmental information not included in this document.

## Waste Electrical and Electronic Equipment (WEEE)



**EU Customers** At the end of the product life cycle, all NI products must be disposed of according to local laws and regulations. For more information about how to recycle NI products in your region, visit [ni.com/environment/weee](http://ni.com/environment/weee).

## 电子信息产品污染控制管理办法（中国 RoHS）



**中国客户** National Instruments 符合中国电子信息产品中限制使用某些有害物质指令 (RoHS)。关于 National Instruments 中国 RoHS 合规性信息，请登录 [ni.com/environment/rohs\\_china](http://ni.com/environment/rohs_china)。(For information about China RoHS compliance, go to [ni.com/environment/rohs\\_china](http://ni.com/environment/rohs_china).)

## Calibration

You can obtain the calibration certificate and information about calibration services for the NI 9252 at [ni.com/calibration](http://ni.com/calibration).

Calibration interval	2 years
----------------------	---------

---

## Appendix

---

### NI 9252 Filtering

The NI 9252 supports two types of lowpass filtering:

- Butterworth
- Comb

**Table 6.** Comparing NI 9252 Filters

Attribute	Butterworth	Comb
Passband	Configurable independent of sample rate	Tracks sample rate
Latency	Medium to high (configuration-dependent)	Low
Phase Delay Variation versus Frequency	Variable input delay	Constant input delay
Flatness	Best	Good
Step Response (Time Domain)	Mid-level delay, overshoot	Short delay, no overshoot/undershoot
Typical Applications	Filtering out high frequency noise sources Reducing measurement noise	Filtering out specific noise sources Control applications

Refer to the specifications for details on the amount of variation in the response you can expect for different input frequency ranges.

## Frequency Response of NI 9252 Filters


The NI 9252 uses programmable hardware filtering to provide an accurate representation of in-band signals and reject out-of-band signals. The filters discriminate between signals based on the frequency range, or bandwidth, of the signal. How the filter discriminates signals based on their frequency is known as frequency response. In general, the frequency response of a filter is described by a signal attenuation (magnitude response) and a input delay (phase response) for every input frequency.

- **Magnitude Response**—The three important frequency ranges, or bandwidths, to consider for magnitude response are passband, transition band, and stopband:
  - **Passband**—The range of frequencies at which the filter attempts to pass a signal without modifying it. The small amount of variation in magnitude at these frequencies is called passband flatness. This is the frequency range of signals that you want to measure.
  - **Transition band**—The range of frequencies in which the filter magnitude response has started to roll-off such that it attenuates signals by some amount, but has not reached the full attenuation amount. The shape of the transition band has an impact

on the alias rejection and how signals are represented in the time domain (for example, step response).

- **Stopband**—The range of frequencies at which the filter attenuates input signals to its maximum attenuation level. Ideally, you want to choose a filter with a stopband that covers frequencies of noise sources that you do not want in your measurements.
- **Input Delay**—Filters delay the input signal by some amount when processing data. In some cases, the delay is a function of the input signal frequency; when this is the case, the input delay plot is useful for knowing the exact delay at different input frequencies and the maximum variation between signals of different frequencies within the passband.


**Figure 9.** Comparing Typical Input Delay for NI 9252 Filters



Each NI 9252 filter has a different frequency response to serve different applications:

- **Butterworth**—Has a passband independent of the sampling rate (as opposed to the comb filter), which offers more flexibility when filtering out noise that is below one-half of the sample rate. However, depending on your settings, you may see alias components of higher frequency signals in your measurement that extend beyond one-half of the sample rate due to the larger transition band.
- **Comb**—Has a smaller passband because its transition band starts early in the frequency range. The comb filter has shorter group delay than other filters and better representation of signals in the time domain (step response). The comb filter's transition band features equally-spaced notches at different frequencies. It is common to use the comb filter with a specific sample rate to align the notches of the transition band thereby removing a specific noise-source frequency from measurements.

**Figure 10.** Comparing Typical Magnitude Response of NI 9252 Filters



The NI 9252 filter delay across signals in the passband varies between filters:

- **Butterworth**—Delays signals by a variable amount depending on their frequency.
- **Comb**—All input frequencies have the same amount of delay when going through the filter. Choose this filter for applications where linear phase, short delay, or data correlation of different devices and configurations is required.

Refer to the specifications for details on the amount of variation in the passband gain and input delay you can expect for different input frequency ranges.


## Step Response of NI 9252 Filters

The shape of the magnitude and phase responses of a filter impacts how signals look in the time domain. The step response of a filter is typically used to identify the behavior of a filter in the time domain.

Three important factors of the filter step response are group delay, rise time, and overshoot/undershoot. The three filters differ in step response across signals in the transition band:

- **Butterworth**—Has a short group delay and the longest rise time. The output signal shows overshoot.
- **Comb**—Has the shortest group delay and the shortest rise time. The output signal does not show overshoot or undershoot.

**Figure 11.** Comparing Typical Step Response of NI 9252 Filters



Information is subject to change without notice. Refer to the *NI Trademarks and Logo Guidelines* at [ni.com/trademarks](http://ni.com/trademarks) for information on NI trademarks. Other product and company names mentioned herein are trademarks or trade names of their respective companies. For patents covering NI products/technology, refer to the appropriate location: **Help»Patents** in your software, the `patents.txt` file on your media, or the *National Instruments Patent Notice* at [ni.com/patents](http://ni.com/patents). You can find information about end-user license agreements (EULAs) and third-party legal notices in the `readme` file for your NI product. Refer to the *Export Compliance Information* at [ni.com/legal/export-compliance](http://ni.com/legal/export-compliance) for the NI global trade compliance policy and how to obtain relevant HTS codes, ECCNs, and other import/export data. NI MAKES NO EXPRESS OR IMPLIED WARRANTIES AS TO THE ACCURACY OF THE INFORMATION CONTAINED HEREIN AND SHALL NOT BE LIABLE FOR ANY ERRORS. U.S. Government Customers: The data contained in this manual was developed at private expense and is subject to the applicable limited rights and restricted data rights as set forth in FAR 52.227-14, DFAR 252.227-7014, and DFAR 252.227-7015.

© 2019 National Instruments. All rights reserved.

378058A-02 June 13, 2019